

Inhibitor-Tolerant qPCR Mix

Product Handling Guide

Shipping: On Dry or Blue Ice
Catalog number: MDX013
Batch No.: See vial
Concentration: 2x

Store at -20 °C



Storage and stability:

Inhibitor-Tolerant qPCR Mix is shipped on dry or blue ice. On arrival store at -20 °C for optimum stability. Repeated freeze/thaw cycles should be avoided. Thawing during transportation does not affect the product performance. Solutions should be mixed/equilibrated after each thawing to avoid phasing.

Expiry:

When stored under the recommended conditions and handled correctly, full activity of the kit is retained until the expiry date on the outer box label.

Safety precautions:

Read and understand the SDS (Safety Data Sheets) before handling the reagents. Hardcopies of the SDS will be provided with the first shipment, thereafter they will be available upon request.

Quality control:

Bioline operates under ISO 13485 Quality Management System. Inhibitor-Tolerant qPCR Mix and its components are extensively tested for activity, processivity, efficiency, heat activation sensitivity, absence of nuclease contamination and absence of nucleic acid contamination.

Notes:

This reagent has been manufactured under 13485 Quality Management System and is suitable for further manufacturing use as an IVD component.

Description

Inhibitor-Tolerant qPCR Mix is a combination of the latest advances in buffer chemistry and PCR enhancers and stabilizers, together with an antibody-mediated hot-start polymerase, dNTPs and MgCl₂. Inhibitor-Tolerant qPCR Mix has been designed for highly reproducible, accurate assay results in the presence of inhibitors, making it ideal for direct amplification directly from human and animal blood samples.

Kit components

Table 1

Component
Inhibitor-Tolerant qPCR Mix, 2x

Users Guidelines

Master mix preparation

Recommended reagent volumes per 20 µL qPCR mix are given in Table 2.

Table 2

Reagent	Volume
Inhibitor-Tolerant qPCR Mix, 2x	10 µL
Primer-Probe Mix, 20x	1 µL
Template*	x µL
Water	As required
Total volume	Up to 20 µL

*If working directly with blood, due to the high viscosity, pipette a minimum of 4 µL of undiluted or diluted blood. The maximum recommended final concentration in the reaction could be dependent of the qPCR platform, this may vary between 5 up to 20 %. Some optimization may be required.

Resuspend reactions with care to minimise the formation of bubbles. Centrifuge reaction plate at 1200 x g for 6 minutes at 4 °C.

Assay setup

The qPCR conditions in Table 3 are suitable for amplicons of up to 200 bp. These cycling parameters below are recommended for blood assays on a number of platforms, however they can be varied to suit different machine-specific protocols.

Table 3

Step	Temperature	Time	Cycles
Polymerase activation	95 °C	3 min	1
Denaturation	95 °C	10 s	45
Annealing/Extension	60 °C	30 s	

Given the high risk of cross-contamination of samples, we recommend including a non-template control.

Technical Support

For any technical enquiries, please contact our Technical Support team via email at: mbi.tech@meridianlifescience.com

Bioline Reagents Ltd
UNITED KINGDOM

Tel: +44 (0)20 8830 5300
Fax: +44 (0)20 8452 2822

Bioline USA Inc.
USA

Tel: +1 901 382 8716
Fax: +1 901 382 0027

Bioline GmbH
GERMANY

Tel: +49 (0)337 160222 00
Fax: +49 (0)3371 60222 01

Bioline (Aust) Pty. Ltd
AUSTRALIA

Tel: +61 (0)2 9209 4180
Fax: +61 (0)2 9209 4763

Bioline France
FRANCE

Tel: +33 (0)1 42 56 04 40
Fax: +33 (0)9 70 06 62 10